



Gmicro Testing



201819000883



TESTING
CNAS L1747

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

REPORT FOR ANALYSIS

Report №. 2018FM13404R01D

Name of Sample Deep ultraviolet LED

Applicant Shenzhen Silverlight Technologies Co.Ltd

Test Type Entrustment Test

Address: Building 66, No.100 Central Xian Lie Road, Guangzhou, China

Postcode: 510070

Tel: (020)87137666

Fax: (020)87137668

Website: www.gddcm.com





GUANGDONG DETECTION CENTER OF MICROBIOLOGY

ANALYSIS AND TEST RESULT

(Report No.): 2018FM13404R01D

Action time:1min

Irradiation distance:10mm

Test Result:

Test germ	Test group	Average cfu of positive controls (cfu/piece)	Average cfu of testing groups (cfu/piece)	Killing rate (%)	Sterilization logarithm
(Escherichia coli) 8099	1	4.4×10^6	3.8×10^5	99.91	3.1
	2	3.2×10^6	3.1×10^5	99.90	3.0
	3	3.8×10^6	3.5×10^5	99.91	3.0

(blank below)



(Report №.): 2018FM13404R01D

Notice Items

1. The Test report is invalid if not affixed with Authorized Stamp of Test and Paging Seal.
2. The Test report is invalid without signature of verifier and approver.
3. The Test report is invalid if being supplemented, deleted or altered.
4. Without prior written permission, the report cannot be reproduced, except in full.
5. Unless otherwise stated, the results shown in this test report refer only to the sample(s) submitted.
6. Any dispute of the report must be raised to the testing body within 15 days after the report is received, exceeding which the dispute will not be accepted.
7. For the tested sample(s) submitted by the applicant, the sample information in the test report is provided by the applicant and the laboratory is not responsible for its authenticity.